

BATCH CERTIFICATE

For Research Use Only

PRODUCT INFORMATION AND QUALITY CONTROL

NAME OF PRODUCT	Ashkenazim son gDNA		
DESCRIPTION	Ashkenazim son gDNA is a highly characterized human DNA from cell lines.		
CATALOG NUMBER	SID-000004		
BATCH NUMBER	00781		
MANUFACTURING CONDITIONS	<ul style="list-style-type: none"> · Manufactured and sealed in class 2 safety cabinet · Manufactured according to DIN EN ISO 13485:2016 		
PACKAGE SIZE AND TYPE	<ul style="list-style-type: none"> · 2D barcoded tube with screw cap · Material: Polypropylen (PP) 		
DATE OF MANUFACTURE	14.08.2024		
EXPIRY DATE	13.08.2026		
TARGET CONCENTRATION	50 ng/μl (dsDNA)		
TARGET QUANTITY	5 μg (dsDNA)		
NOMINAL VOLUME	100 μl		
QUALITY	DNA quantity metrologically traceable to internationally certified reference material (ERM_AD442K).		
STORAGE CONDITIONS	+ 2–8 °C		
MANUFACTURING SITE	SensID GmbH Schillingallee 68, 18057 Rostock, Germany		
TEST METHOD AND ACCEPTANCE CRITERIA	Quality control	Test method	Acceptance criteria
	Quality	Agarose gel electrophoresis	Bright band of high-molecular-weight gDNA ≥ 20 kb
	Quantification	dsDNA measurement: Qubit dsDNA BR Assay Kit (Invitrogen)	dsDNA: 50.0 ng/μl (45.0 – 55.0 ng/μl)
RESULTS OF ANALYSIS	Quality control	Result	PASS / FAIL
	Quality		PASS
	Quantification	48.8 ng/μl (dsDNA)	PASS
COMMENTS / REMARKS	not applicable		

Name and position/title of person authorising the batch release:

Björn Nowack, Managing Director

Date of batch release: 19.08.2024

Signature batch release: Björn Nowack

This document has been created electronically and is valid without signature.