

BATCH CERTIFICATE

For Research Use Only

PRODUCT INFORMATION AND QUALITY CONTROL

NAME OF PRODUCT	Plasma 2X (human-tech) DNA-free dilution buffer			
DESCRIPTION	Dilution buffer for Plasma 2X (human-tech) DNA-free			
CATALOG NUMBER	SID-000005			
BATCH NUMBER	00268			
MANUFACTURING CONDITIONS	<ul style="list-style-type: none">· Manufactured und sealed in class 2 safety cabinet· Bottled with qualified Dispenser· At room temperature· Manufactured according to DIN EN ISO 13485:2016			
PACKAGE SIZE AND TYPE	<ul style="list-style-type: none">· 2D barcoded tube with screw cap· Material: Polypropylen (PP)			
DATE OF MANUFACTURE	13.12.2021			
EXPIRY DATE	12.12.2023			
NOMINAL VOLUME	3.0 ml			
QUALITY	DNA-free			
STORAGE CONDITIONS	+ 2-8 °C			
MANUFACTURING SITE	SensID GmbH Schillingallee 68, 18057 Rostock, Germany			
RESULTS OF ANALYSIS	Result	Units	Specification	Test value
	DNA	pg	≤ 0.3	< 0.3 pg
COMMENTS / REMARKS	not applicable			

Name and position/title of person authorising the batch release:

Björn Nowack, Managing Director

Date of batch release: 17.01.2022

Signature batch release: Björn Nowack

This document has been created electronically and is valid without signature.