



CE according to

directive 98/79/EC

Release Date Version / Index Print Date 27.10.2022 2 1 27.10.2022

BATCH CERTIFICATE

PRODUCT INFORMATION AND QUALITY CONTROL

NAME OF PRODUCT	ldylla™ MSI 0 % FFPE Reference Standard		
DESCRIPTION	The Idylla [™] MSI 0% Reference Standard is a negative control for seven new microsatellite instability (MSI) markers in FFPE format containing DNA from human microsatellite stable cell lines.		
CATALOG NUMBER	SID-000116		
BATCH NUMBER	00261		
MANUFACTURING CONDITIONS	 Manufactured according to DIN EN ISO 13485:2016 		
PACKAGE SIZE AND TYPE	 1.5 ml tube with screw cap Material: Polypropylen (PP) 		
DATE OF MANUFACTURE	14.12.2021		
EXPIRY DATE	13.12.2023		
FORMAT	10 μm section / curl		
AVAILABLE MSI MARKERS	The Idylla [™] MSI 0% Reference Standard covers seven new microsatellite instability (MSI) markers: ACVR2A, BTBD7, DIDO1, MRE11, RYR3, SEC31A and SULF2		
BLEND OF MSI CELL LINE	0%		
QUALITY	DNA quantity metrologically traceable to internationally certified reference material (ERM_AD442K).		
STORAGE CONDITIONS	2-8°C		
MANUFACTURING SITE	SensID GmbH Schillingallee 68, 18057 Rostock, Germany		





cc according to directive 98/79/EC Release Date Version / Index Print Date 27.10.2022 2 1 27.10.2022

TEST METHOD AND ACCEPTANCE CRITERIA	Quality control	Test method	Acceptance criteria
	Quantification	dsDNA measurement: Qubit** dsDNA BR Assay Kit (Invitrogen)	dsDNA: ≥ 110 ng
	DNA Quality	Measurement of DNA Integrity Number (DIN): Tape Station Genomic DNA ScreenTape (Agilent Technologies)	DIN: ≥ 5
	Fraction of MSI	Fraction analysis of MSI/MSS cell lines by ddPCR** (BioRad QX200™)	MSI Fraction 0.0 % (0.0-1.0 %)
	**Measured after extraction with Qiagen QIAamp DNA FFPE Tissue Kit		
RESULTS OF ANALYSIS	Quality control	Result	PASS / FAIL
	Quantification	249.3 ng (dsDNA)	PASS
	DNA Quality	6.9	PASS
	Fraction of MSI	0.10	PASS

Name and position/title of person authorising the batch release:

Björn Nowack, Managing Director

Date of batch release:

07.02.2022

Signature batch release:

Björn Nowack

This document has been created electronically and is valid without signature.