BATCH CERTIFICATE

For Research Use Only

PRODUCT INFORMATION AND QUALITY CONTROL

NAME OF PRODUCT	1X Plasma (human-tech) DNA free, 1000 ml Bulk		
DESCRIPTION	Human serum proteins in common plasma concentrations (1-fold concentrated), electrolytes, EDTA.		
CATALOG NUMBER	SID-000151		
BATCH NUMBER	**00940**		
MANUFACTURING CONDITIONS	· Manufactured and sealed in class 2 safety cabinet · Manufactured according to DIN EN ISO 13485:2016		
PACKAGE SIZE AND TYPE	· 1000 ml Nalgene™ Square Media Bottle · Polyethylene Terephthalate (PET)		
DATE OF MANUFACTURE	02.06.2025		
EXPIRY DATE	01.06.2027		
CONCENTRATION	1X		
NOMINAL VOLUME	1000 ml		
QUALITY	DNA-free		
STORAGE CONDITIONS	+ 2-8 °C		
MANUFACTURING SITE	SensID GmbH Schillingallee 68, 18057 Rostock, Germany		
TEST METHOD AND ACCEPTANCE CRITERIA	Quality control	Test method	Acceptance criteria
	Quality (DNA-free) LOD	Real-Time PCR Investigator Quantiplex Pro Kit (Qiagen)	< 0.5 pg/µl
RESULTS OF ANALYSIS		Result	PASS / FAIL
	Quality	< 0.5 pg/µl	PASS
COMMENTS / REMARKS	Starting materials were tested and found non-reactive for HbsAG and negative for anti-HIV 1 and 2, anti-HCV and HCV RNA.		

Name and position/title of person authorising the batch release:

Björn Nowack, Managing Director

Date of batch release: 13.06.2027

Signature batch release: Björn Nowack

This document has been created electronically and is valid without signature.